

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042923\
 Data File : VN077521.D
 Acq On : 29 Apr 2023 16:04
 Operator : JC\MD
 Sample : PB152434ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB152434ZHE#05

Quant Time: May 01 01:24:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	680206	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1294702	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	1200492	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	422584	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	461371	51.980	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 103.960%		
35) Dibromofluoromethane	8.172	113	395852	50.870	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.740%		
50) Toluene-d8	10.572	98	1595132	53.485	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 106.960%		
62) 4-Bromofluorobenzene	12.854	95	511326	52.402	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 104.800%		
Target Compounds						
					Qvalue	
16) Acetone	4.443	43	83895	29.710	ug/l #	83
20) Methylene Chloride	5.295	84	40065	4.474	ug/l	90
37) Ethyl Acetate	7.572	43	102160	11.958	ug/l	97
43) Isopropyl Acetate	8.695	43	572988	38.770	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042923\
Data File : VN077521.D
Acq On : 29 Apr 2023 16:04
Operator : JC\MD
Sample : PB152434ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB152434ZHE#05

Quant Time: May 01 01:24:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

