

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042923\
 Data File : VN077524.D
 Acq On : 29 Apr 2023 17:17
 Operator : JC\MD
 Sample : PB152434ZHE#08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB152434ZHE#08

Quant Time: May 01 01:24:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	591665	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1132831	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	1052561	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	379504	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	400218	51.838	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 103.680%	
35) Dibromofluoromethane	8.172	113	350096	51.418	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.840%	
50) Toluene-d8	10.566	98	1404322	53.815	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 107.640%	
62) 4-Bromofluorobenzene	12.854	95	448812	52.568	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.140%	
Target Compounds						
						Qvalue
11) Tert butyl alcohol	5.519	59	7738m	6.243	ug/l	
16) Acetone	4.437	43	45520	18.533	ug/l	97
18) Methyl Acetate	5.043	43	10197	1.075	ug/l #	61
20) Methylene Chloride	5.290	84	25600	3.287	ug/l #	81
25) 2-Butanone	7.495	43	32848	9.429	ug/l #	80
37) Ethyl Acetate	7.572	43	90314	12.082	ug/l	97
43) Isopropyl Acetate	8.689	43	502835	38.885	ug/l #	85
52) Toluene	10.630	92	19956	1.017	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042923\
Data File : VN077524.D
Acq On : 29 Apr 2023 17:17
Operator : JC\MD
Sample : PB152434ZHE#08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB152434ZHE#08

Quant Time: May 01 01:24:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

