

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042923\
 Data File : VN077525.D
 Acq On : 29 Apr 2023 17:41
 Operator : JC\MD
 Sample : 02502-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BU-511-VOC-01

Quant Time: May 01 01:25:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	676595	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1284004	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	1165860	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	403188	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	456509	51.707	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 103.420%	
35) Dibromofluoromethane	8.172	113	386929	50.137	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.280%	
50) Toluene-d8	10.572	98	1568733	53.038	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 106.080%	
62) 4-Bromofluorobenzene	12.854	95	490413	50.678	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.360%	
Target Compounds						
						Qvalue
16) Acetone	4.437	43	50080	17.830	ug/l	93
18) Methyl Acetate	5.042	43	11225	1.035	ug/l #	79
20) Methylene Chloride	5.284	84	30945	3.474	ug/l #	82
37) Ethyl Acetate	7.566	43	83927	9.906	ug/l	98
43) Isopropyl Acetate	8.695	43	461043	31.455	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042923\
Data File : VN077525.D
Acq On : 29 Apr 2023 17:41
Operator : JC\MD
Sample : 02502-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BU-511-VOC-01

Quant Time: May 01 01:25:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

