

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042923\
 Data File : VN077528.D
 Acq On : 29 Apr 2023 18:54
 Operator : JC\MD
 Sample : 02503-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-305A-VOC-02

Quant Time: May 01 01:25:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	567481	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1084662	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	1018681	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	347862	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	383111	51.737	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 103.480%	
35) Dibromofluoromethane	8.172	113	332561	51.012	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.020%	
50) Toluene-d8	10.572	98	1337374	53.526	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 107.060%	
62) 4-Bromofluorobenzene	12.854	95	423350	51.788	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.580%	
Target Compounds						
					Qvalue	
16) Acetone	4.437	43	52847	22.433	ug/l #	86
18) Methyl Acetate	5.054	43	9607	1.056	ug/l #	82
20) Methylene Chloride	5.295	84	27889	3.733	ug/l	91
37) Ethyl Acetate	7.572	43	80440	11.239	ug/l	98
43) Isopropyl Acetate	8.695	43	431376	34.840	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042923\
Data File : VN077528.D
Acq On : 29 Apr 2023 18:54
Operator : JC\MD
Sample : 02503-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR-305A-VOC-02

Quant Time: May 01 01:25:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

