

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042923\
 Data File : VN077529.D
 Acq On : 29 Apr 2023 19:18
 Operator : JC\MD
 Sample : 02503-18
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-305A-VOC-03

Quant Time: May 01 01:25:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	561097	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1097401	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	1022594	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	350193	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	380697	51.996	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.000%
35) Dibromofluoromethane	8.172	113	336357	50.995	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.000%
50) Toluene-d8	10.572	98	1340531	53.029	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.060%
62) 4-Bromofluorobenzene	12.854	95	417869	50.524	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.040%
Target Compounds						
						Qvalue
16) Acetone	4.448	43	50744	21.785	ug/l	92
18) Methyl Acetate	5.054	43	10061	1.118	ug/l #	73
20) Methylene Chloride	5.284	84	28479	3.855	ug/l #	80
37) Ethyl Acetate	7.572	43	89888	12.414	ug/l	98
43) Isopropyl Acetate	8.695	43	451655	36.055	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042923\
Data File : VN077529.D
Acq On : 29 Apr 2023 19:18
Operator : JC\MD
Sample : 02503-18
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR-305A-VOC-03

Quant Time: May 01 01:25:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

