

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042923\
 Data File : VN077533.D
 Acq On : 29 Apr 2023 20:55
 Operator : JC\MD
 Sample : 02535-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: May 01 01:26:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	695799	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1299084	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	1134890	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.801	152	394347	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	456952	50.329	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.660%	
35) Dibromofluoromethane	8.172	113	401117	51.372	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.740%	
50) Toluene-d8	10.571	98	1540951	51.494	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.980%	
62) 4-Bromofluorobenzene	12.854	95	473436	48.355	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.720%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042923\
Data File : VN077533.D
Acq On : 29 Apr 2023 20:55
Operator : JC\MD
Sample : 02535-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: May 01 01:26:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

