

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042923\
 Data File : VN077535.D
 Acq On : 29 Apr 2023 21:43
 Operator : JC\MD
 Sample : 02535-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-9

Quant Time: May 01 01:26:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 05/01/2023
 Supervised By : Mahesh Dadoda 05/01/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	593395	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1109484	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	969162	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	337812	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	385129	49.738	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	99.480%		
35) Dibromofluoromethane	8.178	113	345148	51.758	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	103.520%		
50) Toluene-d8	10.572	98	1322698	51.754	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	103.500%		
62) 4-Bromofluorobenzene	12.854	95	400825	47.935	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	95.880%		
Target Compounds						
11) Tert butyl alcohol	5.537	59	19618	15.782	ug/l #	72
16) Acetone	4.448	43	8475m	3.440	ug/l	
18) Methyl Acetate	5.054	43	6357	0.668	ug/l	95
19) Methyl tert-butyl Ether	5.807	73	14125619	641.408	ug/l	99
22) Diisopropyl ether	6.684	45	19897	1.178	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042923\
Data File : VN077535.D
Acq On : 29 Apr 2023 21:43
Operator : JC\MD
Sample : 02535-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-9

Quant Time: May 01 01:26:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/01/2023
Supervised By :Mahesh Dadoda 05/01/2023

