

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042923\
 Data File : VN077512.D
 Acq On : 29 Apr 2023 11:00
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: May 01 01:22:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	0.598	0.594	0.7	93	0.00
3 P	Chloromethane	0.501	0.432	13.8	89	0.00
4 C	Vinyl Chloride	0.668	0.612	8.4#	92	0.00
5 T	Bromomethane	0.578	0.499	13.7	91	0.00
6 T	Chloroethane	0.507	0.421	17.0	89	0.00
7 T	Trichlorofluoromethane	0.981	0.953	2.9	98	0.00
8 T	Diethyl Ether	0.357	0.342	4.2	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.571	0.561	1.8	101	0.00
10 T	Methyl Iodide	0.701	0.681	2.9	93	0.00
11 T	Tert butyl alcohol	0.105	0.083	21.0	83	0.00
12 CM	1,1-Dichloroethene	0.558	0.520	6.8#	97	0.00
13 T	Acrolein	0.104	0.080	23.1	83	0.00
14 T	Allyl chloride	0.620	0.561	9.5	86	0.00
15 T	Acrylonitrile	0.238	0.220	7.6	94	0.00
16 T	Acetone	0.208	0.252	-21.2	133	0.00
17 T	Carbon Disulfide	1.418	1.279	9.8	92	0.00
18 T	Methyl Acetate	0.802	0.707	11.8	92	0.00
19 T	Methyl tert-butyl Ether	1.856	1.837	1.0	101	0.00
20 T	Methylene Chloride	0.658	0.605	8.1	96	0.00
21 T	trans-1,2-Dichloroethene	0.603	0.590	2.2	101	0.00
22 T	Diisopropyl ether	1.424	1.390	2.4	99	0.00
23 T	Vinyl Acetate	0.889	0.829	6.7	92	0.00
24 P	1,1-Dichloroethane	1.008	1.001	0.7	99	0.00
25 T	2-Butanone	0.294	0.290	1.4	102	0.00
26 T	2,2-Dichloropropane	0.977	1.009	-3.3	103	0.00
27 T	cis-1,2-Dichloroethene	0.715	0.707	1.1	101	0.00
28 T	Bromochloromethane	0.385	0.403	-4.7	96	0.00
29 T	Tetrahydrofuran	0.184	0.160	13.0	89	0.00
30 C	Chloroform	1.136	1.162	-2.3#	104	0.00
31 T	Cyclohexane	0.924	0.820	11.3	95	0.00
32 T	1,1,1-Trichloroethane	1.044	1.070	-2.5	100	0.00
33 S	1,2-Dichloroethane-d4	0.652	0.705	-8.1	109	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	96	0.00
35 S	Dibromofluoromethane	0.301	0.329	-9.3	108	0.00
36 T	1,1-Dichloropropene	0.434	0.447	-3.0	99	0.00
37 T	Ethyl Acetate	0.330	0.299	9.4	91	0.00
38 T	Carbon Tetrachloride	0.479	0.496	-3.5	99	0.00
39 T	Methylcyclohexane	0.572	0.600	-4.9	99	0.00
40 TM	Benzene	1.323	1.361	-2.9	100	0.00
41 T	Methacrylonitrile	0.172	0.167	2.9	95	0.00
42 TM	1,2-Dichloroethane	0.454	0.468	-3.1	101	0.00
43 T	Isopropyl Acetate	0.571	0.534	6.5	90	0.00
44 TM	Trichloroethene	0.339	0.346	-2.1	99	0.00
45 C	1,2-Dichloropropane	0.310	0.326	-5.2#	102	0.00
46 T	Dibromomethane	0.251	0.253	-0.8	100	0.00
47 T	Bromodichloromethane	0.486	0.519	-6.8	103	0.00
48 T	Methyl methacrylate	0.260	0.242	6.9	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042923\
 Data File : VN077512.D
 Acq On : 29 Apr 2023 11:00
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: May 01 01:22:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.006	14.3	92	0.00
50 S	Toluene-d8	1.152	1.312	-13.9	108	0.00
51 T	4-Methyl-2-Pentanone	0.335	0.305	9.0	91	0.00
52 CM	Toluene	0.866	0.912	-5.3#	99	0.00
53 T	t-1,3-Dichloropropene	0.492	0.537	-9.1	100	0.00
54 T	cis-1,3-Dichloropropene	0.536	0.576	-7.5	100	0.00
55 T	1,1,2-Trichloroethane	0.348	0.353	-1.4	98	0.00
56 T	Ethyl methacrylate	0.488	0.503	-3.1	98	0.00
57 T	1,3-Dichloropropane	0.555	0.578	-4.1	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.165	0.152	7.9	81	0.00
59 T	2-Hexanone	0.243	0.235	3.3	96	0.00
60 T	Dibromochloromethane	0.362	0.391	-8.0	100	0.00
61 T	1,2-Dibromoethane	0.356	0.358	-0.6	98	0.00
62 S	4-Bromofluorobenzene	0.377	0.461	-22.3	113	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.299	0.303	-1.3	99	0.00
65 PM	Chlorobenzene	1.058	1.064	-0.6	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.385	0.384	0.3	100	0.00
67 C	Ethyl Benzene	1.863	1.958	-5.1#	102	0.00
68 T	m/p-Xylenes	0.724	0.756	-4.4	101	0.00
69 T	o-Xylene	0.712	0.741	-4.1	100	0.00
70 T	Styrene	1.114	1.239	-11.2	104	0.00
71 P	Bromoform	0.252	0.266	-5.6	100	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
73 T	Isopropylbenzene	4.640	4.646	-0.1	102	0.00
74 T	N-amyl acetate	1.374	1.155	15.9	89	0.00
75 P	1,1,2,2-Tetrachloroethane	1.458	1.266	13.2	96	0.00
76 T	1,2,3-Trichloropropane	1.163	1.030	11.4	95	0.00
77 T	Bromobenzene	1.011	0.953	5.7	101	0.00
78 T	n-propylbenzene	5.130	5.334	-4.0	103	0.00
79 T	2-Chlorotoluene	3.163	3.167	-0.1	102	0.00
80 T	1,3,5-Trimethylbenzene	3.446	3.844	-11.5	103	0.00
81 T	trans-1,4-Dichloro-2-butene	0.412	0.393	4.6	95	0.00
82 T	4-Chlorotoluene	2.956	3.007	-1.7	101	0.00
83 T	tert-Butylbenzene	3.346	3.328	0.5	103	0.00
84 T	1,2,4-Trimethylbenzene	3.218	3.764	-17.0	102	0.00
85 T	sec-Butylbenzene	4.593	4.870	-6.0	104	0.00
86 T	p-Isopropyltoluene	3.534	3.920	-10.9	104	0.00
87 T	1,3-Dichlorobenzene	1.798	1.813	-0.8	104	0.00
88 T	1,4-Dichlorobenzene	1.776	1.765	0.6	101	0.00
89 T	n-Butylbenzene	2.735	3.236	-18.3	102	0.00
90 T	Hexachloroethane	0.781	0.747	4.4	99	0.00
91 T	1,2-Dichlorobenzene	1.784	1.763	1.2	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.270	0.228	15.6	92	0.00
93 T	1,2,4-Trichlorobenzene	0.475	0.531	-11.8	88	0.02
94 T	Hexachlorobutadiene	0.360	0.377	-4.7	105	0.04
95 T	Naphthalene	1.725	1.958	-13.5	86	0.05

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042923\
Data File : VN077512.D
Acq On : 29 Apr 2023 11:00
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: May 01 01:22:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.467	0.490	-4.9	87	0.09

(#) = Out of Range SPCC's out = 0 CCC's out = 6