



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042923\
Data File : VN077511.D
Acq On : 29 Apr 2023 10:27
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Title : SW846 8260
Last Update : Tue Apr 25 02:50:59 2023



AutoFind: Scans 1854, 1855, 1856; Background Corrected with Scan 1846

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.8	28624	PASS
75	95	30	60	52.4	89448	PASS
95	95	100	100	100.0	170624	PASS
96	95	5	9	6.7	11377	PASS
173	174	0.00	2	0.9	1011	PASS
174	95	50	100	65.6	111931	PASS
175	174	5	9	7.6	8524	PASS
176	174	95	101	95.6	107000	PASS
177	176	5	9	6.6	7045	PASS

Average of 12.842 to 12.854 min.: VN077511.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
36.05	1225	48.00	1098	61.00	8402	72.00	1061
37.05	6358	49.00	6726	61.95	8623	73.00	806
38.05	6030	50.00	28624	63.00	6795	74.00	3134
39.00	2667	51.00	8767	63.80	268	75.00	8944
39.90	158	52.05	430	64.15	361	76.05	7818
43.00	55	54.85	331	64.85	356	76.95	1208
43.20	91	55.20	101	65.10	191	78.00	654
44.00	1393	56.05	3245	66.95	532	78.95	8218
45.00	1822	56.95	5332	68.00	18412	79.95	2471
46.10	107	57.90	219	69.00	17244	80.95	8724
46.85	1459	60.00	1352	69.95	1875	81.90	1490

Average of 12.842 to 12.854 min.: VN077511.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
83.00	51	96.00	11377	110.90	278	119.00	582
85.90	218	96.95	507	112.05	149	124.70	59
86.20	241	102.80	112	112.70	58	125.70	51
87.00	8066	103.90	1063	112.95	195	127.80	158
87.95	5966	104.80	309	114.95	405	128.00	477
90.80	280	105.85	916	115.70	221	128.95	462
90.95	572	106.10	311	115.95	607	129.80	142
92.00	6178	106.90	147	117.00	1536	130.00	733
93.05	7587	107.10	109	117.80	155	130.70	77
94.05	23072	109.85	144	118.05	344	131.00	159
95.05	170624	110.60	60	118.80	526	134.75	244

Average of 12.842 to 12.854 min.: VN077511.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
135.00	158	145.80	128	171.60	184	281.05	377
135.90	62	146.00	104	171.95	589	281.40	52
136.70	106	147.85	336	172.90	1011		
137.00	338	148.10	66	174.00	111931		
138.90	65	149.90	97	175.00	8524		
140.80	393	152.85	215	176.00	107000		
141.00	1603	154.70	95	177.05	7045		
141.90	269	155.05	284	177.90	287		
142.70	376	157.00	531	192.80	59		
143.00	1926	159.00	105	193.10	65		
144.10	71	160.90	64	207.00	174		