

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN042924\  
 Data File : VN081915.D  
 Acq On : 29 Apr 2024 15:32  
 Operator : JC\MD  
 Sample : IBLK  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 15 Sample Multiplier: 1

**Instrument :**  
 MSVOA\_N  
**ClientSampleId :**  
 IBLK

Quant Time: Apr 30 04:38:59 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N042624W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Sat Apr 27 04:37:27 2024  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Bromochloromethane	7.812	128	41833	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.106	114	238539	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.871	117	204159	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.583	65	99128	28.302	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	94.333%
60) 4-Bromofluorobenzene	12.853	95	96875	28.512	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	95.033%
63) Toluene-d8	10.571	98	294926	30.419	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	101.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN042924\  
Data File : VN081915.D  
Acq On : 29 Apr 2024 15:32  
Operator : JC\MD  
Sample : IBLK  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 15 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
IBLK

Quant Time: Apr 30 04:38:59 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N042624W.M  
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
QLast Update : Sat Apr 27 04:37:27 2024  
Response via : Initial Calibration

