

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN043019\
 Data File : VN055300.D
 Acq On : 30 Apr 2019 11:48
 Operator : JC/SP
 Sample : VN0430WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0430WBL01

Quant Time: May 01 08:53:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:16:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	346738	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	590868	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	546951	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	141551	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	242881	53.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.08%	
35) Dibromofluoromethane	7.59	113	196715	50.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.44%	
50) Toluene-d8	10.09	98	732567	52.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.62%	
62) 4-Bromofluorobenzene	12.40	95	223030	50.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.92%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN043019\
Data File : VN055300.D
Acq On : 30 Apr 2019 11:48
Operator : JC/SP
Sample : VN0430WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0430WBL01

Quant Time: May 01 08:53:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration

