

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN043019\
 Data File : VN055310.D
 Acq On : 30 Apr 2019 16:02
 Operator : JC/SP
 Sample : K2194-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JC-03-042919-G

Quant Time: May 01 09:08:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:16:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	364043	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	623132	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	577850	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	178317	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	258762	54.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.66%	
35) Dibromofluoromethane	7.59	113	207187	50.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.32%	
50) Toluene-d8	10.09	98	791253	53.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.16%	
62) 4-Bromofluorobenzene	12.40	95	250551	54.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.56%	
Target Compounds						
16) Acetone	3.82	43	22968	15.478	ug/l	97
20) Methylene Chloride	4.55	84	1051252	232.092	ug/l	98
43) Isopropyl Acetate	8.18	43	20452	2.782	ug/l #	88
95) Naphthalene	15.13	128	1780	3.664	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN043019\
Data File : VN055310.D
Acq On : 30 Apr 2019 16:02
Operator : JC/SP
Sample : K2194-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JC-03-042919-G

Quant Time: May 01 09:08:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration

