

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN043019\
 Data File : VN055311.D
 Acq On : 30 Apr 2019 16:27
 Operator : JC/SP
 Sample : K2194-16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JC-03-042919-H

Quant Time: May 01 09:09:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:16:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	353452	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	609243	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	574439	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	167167	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	251745	54.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.88%	
35) Dibromofluoromethane	7.59	113	201579	50.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.82%	
50) Toluene-d8	10.09	98	765738	53.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.06%	
62) 4-Bromofluorobenzene	12.40	95	243800	54.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.04%	
Target Compounds						
16) Acetone	3.82	43	23298	16.171	ug/l	95
20) Methylene Chloride	4.55	84	29080	6.613	ug/l	97
43) Isopropyl Acetate	8.17	43	22256	3.096	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN043019\
Data File : VN055311.D
Acq On : 30 Apr 2019 16:27
Operator : JC/SP
Sample : K2194-16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JC-03-042919-H

Quant Time: May 01 09:09:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration

