

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN043019\
 Data File : VN055313.D
 Acq On : 30 Apr 2019 17:18
 Operator : JC/SP
 Sample : K2199-18
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-02-042919-A

Quant Time: May 01 09:12:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:16:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	355501	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	630766	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	576383	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	179579	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	254317	54.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.36%	
35) Dibromofluoromethane	7.59	113	205091	49.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.08%	
50) Toluene-d8	10.09	98	773295	51.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.46%	
62) 4-Bromofluorobenzene	12.40	95	248119	53.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.20%	
Target Compounds						
16) Acetone	3.82	43	21973	15.163	ug/l	99
20) Methylene Chloride	4.55	84	173373	39.197	ug/l	98
43) Isopropyl Acetate	8.18	43	23321	3.134	ug/l #	81
95) Naphthalene	15.13	128	34192	9.722	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN043019\
Data File : VN055313.D
Acq On : 30 Apr 2019 17:18
Operator : JC/SP
Sample : K2199-18
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-02-042919-A

Quant Time: May 01 09:12:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration

