

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN043020\
 Data File : VN061200.D
 Acq On : 30 Apr 2020 14:28
 Operator : JC/MD
 Sample : VN0430WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0430WBL01

Quant Time: May 01 03:37:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 29 08:14:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	96679	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	168726	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	156908	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	61341	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.98	65	74837	48.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.84%	
35) Dibromofluoromethane	7.54	113	50727	50.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.68%	
50) Toluene-d8	10.06	98	208922	51.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.74%	
62) 4-Bromofluorobenzene	12.37	95	72120	47.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.24%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN043020\
Data File : VN061200.D
Acq On : 30 Apr 2020 14:28
Operator : JC/MD
Sample : VN0430WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0430WBL01

Quant Time: May 01 03:37:34 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 29 08:14:26 2020
Response via : Initial Calibration

