

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN043021\
 Data File : VN066904.D
 Acq On : 30 Apr 2021 12:37
 Operator : JC/MD
 Sample : VN0430MBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0430MBL01

Quant Time: May 01 01:45:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.588	168	304961	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.516	114	451099	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.348	117	378047	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.293	152	152238	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.950	65	165348	42.20	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	84.40%		
35) Dibromofluoromethane	7.511	113	153137	55.00	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	110.00%		
50) Toluene-d8	10.029	98	572183	52.35	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	104.70%		
62) 4-Bromofluorobenzene	12.349	95	160521	43.41	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	86.82%		
Target Compounds					Qvalue	
95) Naphthalene	15.084	128	32495	4.81	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN043021\
Data File : VN066904.D
Acq On : 30 Apr 2021 12:37
Operator : JC/MD
Sample : VN0430MBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0430MBL01

Quant Time: May 01 01:45:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

