

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN043021\
 Data File : VN066908.D
 Acq On : 30 Apr 2021 14:30
 Operator : JC/MD
 Sample : M2204-05 200X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ANTIFREEZE-1906

Quant Time: May 01 01:46:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	335794	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	492754	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.351	117	430261	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.293	152	183695	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	182809	42.37	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	84.74%
35) Dibromofluoromethane	7.511	113	166768	54.83	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	109.66%
50) Toluene-d8	10.029	98	629414	52.72	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.44%
62) 4-Bromofluorobenzene	12.349	95	187724	46.47	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.94%
Target Compounds						
					Qvalue	
16) Acetone	3.753	43	155061	95.26	ug/l	95
40) Benzene	7.969	78	4789	0.34	ug/l	98
52) Toluene	10.096	92	14031	1.64	ug/l	99
84) 1,2,4-Trimethylbenzene	12.998	105	3048	0.28	ug/l	95
95) Naphthalene	15.087	128	24989	3.06	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN043021\
Data File : VN066908.D
Acq On : 30 Apr 2021 14:30
Operator : JC/MD
Sample : M2204-05 200X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ANTIFREEZE-1906

Quant Time: May 01 01:46:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

