

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN043021\
 Data File : VN066911.D
 Acq On : 30 Apr 2021 15:42
 Operator : JC/MD
 Sample : M2204-05 100X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ANTIFREEZE-1906

Quant Time: May 01 01:46:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	369781	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	547445	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.351	117	475441	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.293	152	199693	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	202630	42.65	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	85.30%
35) Dibromofluoromethane	7.513	113	186502	55.19	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	110.38%
50) Toluene-d8	10.029	98	693053	52.25	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.50%
62) 4-Bromofluorobenzene	12.349	95	213312	47.53	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.06%
Target Compounds						
16) Acetone	3.753	43	301769	168.36	ug/l	95
40) Benzene	7.969	78	8258	0.52	ug/l	96
52) Toluene	10.093	92	25563	2.69	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN043021\
Data File : VN066911.D
Acq On : 30 Apr 2021 15:42
Operator : JC/MD
Sample : M2204-05 100X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ANTIFREEZE-1906

Quant Time: May 01 01:46:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

