

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN043021\
 Data File : VN066912.D
 Acq On : 30 Apr 2021 16:06
 Operator : JC/MD
 Sample : PB136080TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB136080TB

Quant Time: May 01 01:46:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	350411	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	529776	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.348	117	482832	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.293	152	184998	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	198934	44.19	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	88.38%		
35) Dibromofluoromethane	7.513	113	179282	54.82	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	109.64%		
50) Toluene-d8	10.029	98	693849	54.05	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	108.10%		
62) 4-Bromofluorobenzene	12.349	95	209510	48.24	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	96.48%		
Target Compounds						Qvalue
16) Acetone	3.761	43	17168	10.11	ug/l	93
18) Methyl Acetate	4.271	43	2120m	1.57	ug/l	
20) Methylene Chloride	4.475	84	17155	3.99	ug/l	93
43) Isopropyl Acetate	8.098	43	23669	2.85	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN043021\
Data File : VN066912.D
Acq On : 30 Apr 2021 16:06
Operator : JC/MD
Sample : PB136080TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB136080TB

Quant Time: May 01 01:46:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

