

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN043021\
 Data File : VN066913.D
 Acq On : 30 Apr 2021 16:30
 Operator : JC/MD
 Sample : M2222-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-5-COMP

Manual Integrations
 APPROVED

MMDadoda
 5/3/2021 11:38:59 AM

Quant Time: May 01 01:47:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	315130	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	473010	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.351	117	442866	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.290	152	186501	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.956	65	177249	43.78	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	87.56%		
35) Dibromofluoromethane	7.516	113	162209	55.55	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	111.10%		
50) Toluene-d8	10.029	98	626288	54.65	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	109.30%		
62) 4-Bromofluorobenzene	12.349	95	195032	50.30	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	100.60%		
Target Compounds						Qvalue
16) Acetone	3.761	43	14894	9.75	ug/l	98
18) Methyl Acetate	4.257	43	2783m	1.77	ug/l	
20) Methylene Chloride	4.472	84	14390	3.72	ug/l	92
43) Isopropyl Acetate	8.095	43	22790	3.07	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN043021\
Data File : VN066913.D
Acq On : 30 Apr 2021 16:30
Operator : JC/MD
Sample : M2222-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-5-COMP

Manual Integrations
APPROVED

MMDadoda
5/3/2021 11:38:59 AM

Quant Time: May 01 01:47:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

