

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN043021\
 Data File : VN066914.D
 Acq On : 30 Apr 2021 16:54
 Operator : JC/MD
 Sample : M2229-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-6

Manual Integrations
 APPROVED

MMDadoda
 5/3/2021 11:39:02 AM

Quant Time: May 01 01:47:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	316866	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	472123	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.348	117	434016	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.293	152	178098	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	174836	42.94	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	85.88%		
35) Dibromofluoromethane	7.513	113	161359	55.37	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	110.74%		
50) Toluene-d8	10.029	98	622471	54.42	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	108.84%		
62) 4-Bromofluorobenzene	12.349	95	192847	49.83	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	99.66%		
Target Compounds					Qvalue	
16) Acetone	3.756	43	29524	19.22	ug/l	95
18) Methyl Acetate	4.265	43	3147m	1.86	ug/l	
20) Methylene Chloride	4.474	84	19065	4.90	ug/l	90
43) Isopropyl Acetate	8.098	43	22280	3.01	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN043021\
Data File : VN066914.D
Acq On : 30 Apr 2021 16:54
Operator : JC/MD
Sample : M2229-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-6

Manual Integrations
APPROVED

MMDadoda
5/3/2021 11:39:02 AM

Quant Time: May 01 01:47:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

