

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN043024\
 Data File : VN081923.D
 Acq On : 30 Apr 2024 11:56
 Operator : JC\MD
 Sample : P2183-03 20X
 Misc : 1.10g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1232

Quant Time: May 01 00:50:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	212568	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	413644	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	374034	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	157037	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	158168	44.088	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.180%
35) Dibromofluoromethane	8.171	113	121488	48.781	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.560%
50) Toluene-d8	10.565	98	513439	52.492	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.980%
62) 4-Bromofluorobenzene	12.847	95	180229	46.612	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	93.220%
Target Compounds						
					Qvalue	
16) Acetone	4.418	43	266795	251.516	ug/l	98
52) Toluene	10.629	92	293003	40.226	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN043024\
Data File : VN081923.D
Acq On : 30 Apr 2024 11:56
Operator : JC\MD
Sample : P2183-03 20X
Misc : 1.10g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1232

Quant Time: May 01 00:50:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration

