

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN043024\
 Data File : VN081935.D
 Acq On : 30 Apr 2024 16:53
 Operator : JC\MD
 Sample : P2307-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-2

Quant Time: May 01 00:56:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	206127	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	412638	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	367576	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	146086	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	157242	45.199	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.400%
35) Dibromofluoromethane	8.165	113	118204	47.578	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.160%
50) Toluene-d8	10.565	98	506996	51.960	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.920%
62) 4-Bromofluorobenzene	12.847	95	174062	45.127	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	90.260%
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	52345	46.431	ug/l	98
20) Methylene Chloride	5.277	84	12314	4.130	ug/l	89
43) Isopropyl Acetate	8.689	43	379007	46.124	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN043024\
Data File : VN081935.D
Acq On : 30 Apr 2024 16:53
Operator : JC\MD
Sample : P2307-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-2

Quant Time: May 01 00:56:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration

