

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN043024\
 Data File : VN081936.D
 Acq On : 30 Apr 2024 17:17
 Operator : JC\MD
 Sample : P2307-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-3

Quant Time: May 01 00:56:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	230910	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	453649	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	400798	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	156653	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	174030	44.656	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.320%
35) Dibromofluoromethane	8.171	113	133594	48.911	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.820%
50) Toluene-d8	10.571	98	544802	50.787	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.580%
62) 4-Bromofluorobenzene	12.853	95	185107	43.652	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	87.300%
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	55833	43.942	ug/l	97
20) Methylene Chloride	5.283	84	13306	3.984	ug/l #	87
43) Isopropyl Acetate	8.688	43	327410	36.243	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN043024\
Data File : VN081936.D
Acq On : 30 Apr 2024 17:17
Operator : JC\MD
Sample : P2307-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-3

Quant Time: May 01 00:56:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration

