

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN043024\
 Data File : VN081939.D
 Acq On : 30 Apr 2024 18:29
 Operator : JC\MD
 Sample : P2183-01DL 100X
 Misc : 5.03g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1229DL

Quant Time: May 01 06:02:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	215931	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	418692	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	368919	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	148604	50.000	ug/l	# 0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	154449	42.380	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	84.760%
35) Dibromofluoromethane	8.165	113	122766	48.700	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.400%
50) Toluene-d8	10.571	98	514144	51.931	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.860%
62) 4-Bromofluorobenzene	12.847	95	179934	45.975	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	91.940%
Target Compounds						
					Qvalue	
16) Acetone	4.418	43	138184	125.502	ug/l	98
52) Toluene	10.629	92	115904	15.720	ug/l	98
68) m/p-Xylenes	12.071	106	15019	2.810	ug/l	91
69) o-Xylene	12.400	106	5216	1.000	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN043024\
Data File : VN081939.D
Acq On : 30 Apr 2024 18:29
Operator : JC\MD
Sample : P2183-01DL 100X
Misc : 5.03g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1229DL

Quant Time: May 01 06:02:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration

