

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN043025\
 Data File : VN086430.D
 Acq On : 30 Apr 2025 12:37
 Operator : JC\MD
 Sample : PB167775TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167775TB

Quant Time: May 01 01:33:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	156151	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	301708	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	289830	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	123592	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	112342	49.614	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.220%
35) Dibromofluoromethane	8.165	113	72553	51.814	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.620%
50) Toluene-d8	10.565	98	379859	50.758	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.520%
62) 4-Bromofluorobenzene	12.847	95	138835	50.862	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.720%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.518	59	8162	19.756	ug/l	96
16) Acetone	4.436	43	37429	43.086	ug/l	94
20) Methylene Chloride	5.271	84	8380	4.003	ug/l #	83
25) 2-Butanone	7.482	43	38495	27.312	ug/l	98
43) Isopropyl Acetate	8.682	43	33464	4.549	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN043025\
Data File : VN086430.D
Acq On : 30 Apr 2025 12:37
Operator : JC\MD
Sample : PB167775TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167775TB

Quant Time: May 01 01:33:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

