

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN043025\
 Data File : VN086433.D
 Acq On : 30 Apr 2025 13:50
 Operator : JC\MD
 Sample : PB167775ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167775ZHE#03

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/02/2025
 Supervised By : Semsettin Yesilyurt 05/05/2025

Quant Time: May 01 01:34:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	155440	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	291732	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	272591	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	112524	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	113675	50.432	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 100.860%			
35) Dibromofluoromethane	8.165	113	71664	52.929	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 105.860%			
50) Toluene-d8	10.565	98	373699	51.643	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 103.280%			
62) 4-Bromofluorobenzene	12.847	95	127570	48.334	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 96.660%			
Target Compounds						
11) Tert butyl alcohol	5.518	59	7620	18.528	ug/l #	91
16) Acetone	4.436	43	35162	40.333	ug/l	97
20) Methylene Chloride	5.283	84	5474	2.627	ug/l #	89
25) 2-Butanone	7.482	43	35524	25.320	ug/l	97
43) Isopropyl Acetate	8.682	43	34773m	5.030	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN043025\
Data File : VN086433.D
Acq On : 30 Apr 2025 13:50
Operator : JC\MD
Sample : PB167775ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167775ZHE#03

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/02/2025
Supervised By :Semsettin Yesilyurt 05/05/2025

Quant Time: May 01 01:34:21 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

