

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN043025\
 Data File : VN086425.D
 Acq On : 30 Apr 2025 10:17
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: May 01 01:32:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	81	0.00
2 T	Dichlorodifluoromethane	0.593	0.596	-0.5	82	0.00
3 P	Chloromethane	0.861	0.817	5.1	88	0.00
4 C	Vinyl Chloride	0.822	0.798	2.9#	82	0.00
5 T	Bromomethane	0.370	0.432	-16.8	99	0.00
6 T	Chloroethane	0.550	0.530	3.6	86	0.00
7 T	Trichlorofluoromethane	0.911	0.922	-1.2	87	0.00
8 T	Diethyl Ether	0.398	0.378	5.0	82	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.551	0.569	-3.3	88	0.00
10 T	Methyl Iodide	0.606	0.738	-21.8	100	0.00
11 T	Tert butyl alcohol	0.132	0.116	12.1	76	0.00
12 CM	1,1-Dichloroethene	0.588	0.575	2.2#	85	0.00
13 T	Acrolein	0.075	0.106	-41.3#	142	0.00
14 T	Allyl chloride	1.046	0.897	14.2	79	0.00
15 T	Acrylonitrile	0.327	0.310	5.2	80	0.00
16 T	Acetone	0.290	0.243	16.2	79	0.00
17 T	Carbon Disulfide	1.762	1.578	10.4	82	0.00
18 T	Methyl Acetate	0.870	0.699	19.7	73	0.00
19 T	Methyl tert-butyl Ether	2.163	2.073	4.2	84	0.00
20 T	Methylene Chloride	0.670	0.656	2.1	86	0.00
21 T	trans-1,2-Dichloroethene	0.615	0.617	-0.3	87	0.00
22 T	Diisopropyl ether	2.276	2.113	7.2	82	0.00
23 T	Vinyl Acetate	1.591	1.524	4.2	84	0.00
24 P	1,1-Dichloroethane	1.201	1.159	3.5	83	0.00
25 T	2-Butanone	0.451	0.411	8.9	79	0.00
26 T	2,2-Dichloropropane	1.075	1.062	1.2	88	0.00
27 T	cis-1,2-Dichloroethene	0.764	0.756	1.0	86	0.00
28 T	Bromochloromethane	0.509	0.464	8.8	76	0.00
29 T	Tetrahydrofuran	0.302	0.273	9.6	79	0.00
30 C	Chloroform	1.180	1.183	-0.3#	87	0.00
31 T	Cyclohexane	1.175	1.069	9.0	82	0.00
32 T	1,1,1-Trichloroethane	1.009	1.028	-1.9	88	0.00
33 S	1,2-Dichloroethane-d4	0.725	0.716	1.2	81	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	79	0.00
35 S	Dibromofluoromethane	0.232	0.232	0.0	86	0.00
36 T	1,1-Dichloropropene	0.463	0.485	-4.8	87	0.00
37 T	Ethyl Acetate	0.488	0.455	6.8	81	0.00
38 T	Carbon Tetrachloride	0.441	0.473	-7.3	89	0.00
39 T	Methylcyclohexane	0.551	0.557	-1.1	84	0.00
40 TM	Benzene	1.485	1.512	-1.8	85	0.00
41 T	Methacrylonitrile	0.283	0.255	9.9	77	0.00
42 TM	1,2-Dichloroethane	0.474	0.493	-4.0	88	0.00
43 T	Isopropyl Acetate	1.177	0.867	26.3#	79	0.00
44 TM	Trichloroethene	0.352	0.376	-6.8	88	0.00
45 C	1,2-Dichloropropane	0.364	0.360	1.1#	82	0.00
46 T	Dibromomethane	0.232	0.250	-7.8	87	0.00
47 T	Bromodichloromethane	0.500	0.523	-4.6	87	0.00
48 T	Methyl methacrylate	0.429	0.397	7.5	80	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN043025\
 Data File : VN086425.D
 Acq On : 30 Apr 2025 10:17
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: May 01 01:32:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.007	0.0	87	0.00
50 S	Toluene-d8	1.240	1.291	-4.1	82	0.00
51 T	4-Methyl-2-Pentanone	0.500	0.485	3.0	80	0.00
52 CM	Toluene	0.927	0.979	-5.6#	87	0.00
53 T	t-1,3-Dichloropropene	0.558	0.594	-6.5	86	0.00
54 T	cis-1,3-Dichloropropene	0.613	0.623	-1.6	86	0.00
55 T	1,1,2-Trichloroethane	0.333	0.352	-5.7	88	0.00
56 T	Ethyl methacrylate	0.613	0.615	-0.3	84	0.00
57 T	1,3-Dichloropropane	0.592	0.617	-4.2	87	0.00
58 T	2-Chloroethyl Vinyl ether	0.291	0.269	7.6	74	0.00
59 T	2-Hexanone	0.371	0.359	3.2	80	0.00
60 T	Dibromochloromethane	0.366	0.396	-8.2	89	0.00
61 T	1,2-Dibromoethane	0.336	0.363	-8.0	89	0.00
62 S	4-Bromofluorobenzene	0.452	0.488	-8.0	84	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	82	0.00
64 T	Tetrachloroethene	0.385	0.406	-5.5	90	0.00
65 PM	Chlorobenzene	1.116	1.150	-3.0	90	0.00
66 T	1,1,1,2-Tetrachloroethane	0.368	0.386	-4.9	91	0.00
67 C	Ethyl Benzene	2.014	2.109	-4.7#	90	0.00
68 T	m/p-Xylenes	0.754	0.809	-7.3	92	0.00
69 T	o-Xylene	0.746	0.793	-6.3	92	0.00
70 T	Styrene	1.243	1.357	-9.2	92	0.00
71 P	Bromoform	0.277	0.300	-8.3	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
73 T	Isopropylbenzene	4.090	4.029	1.5	93	0.00
74 T	N-amyl acetate	2.036	1.738	14.6	84	0.00
75 P	1,1,2,2-Tetrachloroethane	1.176	1.094	7.0	90	0.00
76 T	1,2,3-Trichloropropane	1.180	1.072	9.2	88	0.00
77 T	Bromobenzene	0.906	0.949	-4.7	96	0.00
78 T	n-propylbenzene	4.820	4.830	-0.2	92	0.00
79 T	2-Chlorotoluene	3.016	2.985	1.0	93	0.00
80 T	1,3,5-Trimethylbenzene	3.348	3.359	-0.3	93	0.00
81 T	trans-1,4-Dichloro-2-butene	0.491	0.513	-4.5	95	0.00
82 T	4-Chlorotoluene	2.996	3.005	-0.3	92	0.00
83 T	tert-Butylbenzene	2.901	2.813	3.0	92	0.00
84 T	1,2,4-Trimethylbenzene	3.408	3.413	-0.1	92	0.00
85 T	sec-Butylbenzene	4.023	4.020	0.1	91	0.00
86 T	p-Isopropyltoluene	3.316	3.389	-2.2	92	0.00
87 T	1,3-Dichlorobenzene	1.704	1.773	-4.0	94	0.00
88 T	1,4-Dichlorobenzene	1.716	1.788	-4.2	95	0.00
89 T	n-Butylbenzene	2.939	2.983	-1.5	90	0.00
90 T	Hexachloroethane	0.558	0.562	-0.7	93	0.00
91 T	1,2-Dichlorobenzene	1.652	1.685	-2.0	93	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.237	0.219	7.6	85	0.00
93 T	1,2,4-Trichlorobenzene	0.791	0.779	1.5	87	0.00
94 T	Hexachlorobutadiene	0.299	0.286	4.3	87	0.00
95 T	Naphthalene	2.804	2.551	9.0	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN043025\
Data File : VN086425.D
Acq On : 30 Apr 2025 10:17
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: May 01 01:32:22 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.749	0.717	4.3	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6