

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN043025\
 Data File : VN086427.D
 Acq On : 30 Apr 2025 11:15
 Operator : JC\MD
 Sample : VN0430WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0430WBL01

Quant Time: May 01 01:32:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	168139	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	316611	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	297458	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	114129	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	118438	48.577	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.160%
35) Dibromofluoromethane	8.165	113	86056	58.564	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	117.120%
50) Toluene-d8	10.565	98	410088	52.218	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.440%
62) 4-Bromofluorobenzene	12.847	95	139992	48.872	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.740%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN043025\
Data File : VN086427.D
Acq On : 30 Apr 2025 11:15
Operator : JC\MD
Sample : VN0430WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0430WBL01


Quant Time: May 01 01:32:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. 0.006 min
Lab File: VN086427.D
Acq: 30 Apr 2025 11:15


Instrument :
MSVOA_N
ClientSampleId :
VN0430WBL01





Tgt Ion:168 Resp: 168139
Ion Ratio Lower Upper
168 100
99 62.2 52.5 78.7




#33
1,2-Dichloroethane-d4
Concen: 48.577 ug/l
RT: 8.577 min Scan# 1126
Delta R.T. 0.000 min
Lab File: VN086427.D
Acq: 30 Apr 2025 11:15



Tgt Ion: 65 Resp: 118438
Ion Ratio Lower Upper
65 100
67 52.8 0.0 106.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN086427.D

Acq: 30 Apr 2025 11:15

Instrument :

MSVOA_N

ClientSampleId :

VN0430WBL01





Tgt Ion:114 Resp: 316611

Ion Ratio Lower Upper

114 100

63 20.7 0.0 42.6

88 15.2 0.0 31.8



#35

Dibromofluoromethane

Concen: 58.564 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.000 min

Lab File: VN086427.D

Acq: 30 Apr 2025 11:15




Tgt Ion:113 Resp: 86056


Ion Ratio Lower Upper

113 100

111 101.9 83.4 125.0

192 17.0 13.7 20.5









#50
Toluene-d8
Concen: 52.218 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. 0.000 min
Lab File: VN086427.D
Acq: 30 Apr 2025 11:15




Instrument :
MSVOA_N
ClientSampleId :
VN0430WBL01


Tgt Ion: 98 Resp: 410088
Ion Ratio Lower Upper
98 100
100 65.5 52.5 78.7



#62
4-Bromofluorobenzene
Concen: 48.872 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN086427.D
Acq: 30 Apr 2025 11:15

Tgt Ion: 95 Resp: 139992
Ion Ratio Lower Upper
95 100
174 72.2 0.0 133.4
176 69.6 0.0 129.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN086427.D
Acq: 30 Apr 2025 11:15

Instrument :
MSVOA_N
ClientSampleId :
VN0430WBL01

Tgt Ion:117 Resp: 297458
Ion Ratio Lower Upper
117 100
82 56.9 44.7 67.1
119 32.8 26.4 39.6



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. 0.000 min
Lab File: VN086427.D
Acq: 30 Apr 2025 11:15

Tgt Ion:152 Resp: 114129
Ion Ratio Lower Upper
152 100
115 61.2 31.9 95.9
150 158.5 0.0 352.0

