

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050119\
 Data File : VN055325.D
 Acq On : 1 May 2019 13:27
 Operator : JC/SP
 Sample : K2191-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 P-66-TP

Quant Time: May 02 08:58:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:16:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	364979	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	624547	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	579663	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	164138	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	254495	53.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.60%	
35) Dibromofluoromethane	7.59	113	208934	50.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.94%	
50) Toluene-d8	10.09	98	786180	53.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.22%	
62) 4-Bromofluorobenzene	12.40	95	241795	52.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.54%	
Target Compounds						
16) Acetone	3.82	43	35093	23.588	ug/l	98
20) Methylene Chloride	4.55	84	368779	81.209	ug/l	98
25) 2-Butanone	6.85	43	13963	7.339	ug/l	100
43) Isopropyl Acetate	8.17	43	16975	2.304	ug/l #	83
95) Naphthalene	15.13	128	6023	4.561	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN050119\
Data File : VN055325.D
Acq On : 1 May 2019 13:27
Operator : JC/SP
Sample : K2191-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
P-66-TP

Quant Time: May 02 08:58:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration

