

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050119\
 Data File : VN055327.D
 Acq On : 1 May 2019 14:18
 Operator : JC/SP
 Sample : K2630-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: May 02 03:42:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:16:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	356954	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	613933	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	555361	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	140936	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	246579	52.80	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.60%	
35) Dibromofluoromethane	7.59	113	200381	49.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.46%	
50) Toluene-d8	10.09	98	762370	52.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.78%	
62) 4-Bromofluorobenzene	12.40	95	226684	49.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.70%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN050119\
Data File : VN055327.D
Acq On : 1 May 2019 14:18
Operator : JC/SP
Sample : K2630-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: May 02 03:42:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration

