

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050119\
 Data File : VN055329.D
 Acq On : 1 May 2019 15:09
 Operator : JC/SP
 Sample : K2630-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 025-OUTFALL(BACK-YARD)

Quant Time: May 02 03:47:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:16:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	361457	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	615113	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	562643	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	147633	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	247616	52.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.72%	
35) Dibromofluoromethane	7.59	113	203495	50.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.80%	
50) Toluene-d8	10.09	98	765389	52.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.00%	
62) 4-Bromofluorobenzene	12.40	95	232109	50.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.88%	
Target Compounds						
16) Acetone	3.82	43	19218	13.043	ug/l	100
70) Styrene	11.96	104	134922	12.504	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN050119\
Data File : VN055329.D
Acq On : 1 May 2019 15:09
Operator : JC/SP
Sample : K2630-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
025-OUTFALL(BACK-YARD)

Quant Time: May 02 03:47:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration

