

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050119\
 Data File : VN055333.D
 Acq On : 1 May 2019 17:11
 Operator : JC/SP
 Sample : K2650-01 10X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GROUND-WATER

Quant Time: May 02 08:48:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:16:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	368141	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	616503	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	553725	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	143428	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	250680	52.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.10%	
35) Dibromofluoromethane	7.59	113	205447	50.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.54%	
50) Toluene-d8	10.09	98	770010	52.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.40%	
62) 4-Bromofluorobenzene	12.40	95	227804	49.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.76%	
Target Compounds						
16) Acetone	3.82	43	15160	10.102	ug/l	100
68) m/p-Xylenes	11.62	106	11309	1.621	ug/l	96
69) o-Xylene	11.94	106	5807	0.840	ug/l	88
84) 1,2,4-Trimethylbenzene	13.04	105	6693	0.666	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN050119\
Data File : VN055333.D
Acq On : 1 May 2019 17:11
Operator : JC/SP
Sample : K2650-01 10X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GROUND-WATER

Quant Time: May 02 08:48:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration

