

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN050119\
 Data File : VN055334.D
 Acq On : 1 May 2019 17:36
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: May 02 04:08:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:16:52 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	90	0.00
2 T	Dichlorodifluoromethane	0.613	0.606	1.1	86	0.00
3 P	Chloromethane	0.759	0.709	6.6	87	0.00
4 C	Vinyl Chloride	0.611	0.568	7.0#	88	0.00
5 T	Bromomethane	0.324	0.303	6.5	86	0.00
6 T	Chloroethane	0.331	0.308	6.9	89	0.00
7 T	Trichlorofluoromethane	0.869	0.842	3.1	93	0.00
8 T	Diethyl Ether	0.304	0.298	2.0	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.518	0.495	4.4	90	0.00
10 T	Methyl Iodide	0.729	0.728	0.1	89	0.00
11 T	Tert butyl alcohol	0.037	0.040	-8.1	97	0.00
12 CM	1,1-Dichloroethene	0.494	0.471	4.7#	90	0.00
13 T	Acrolein	0.029	0.029	0.0	85	0.00
14 T	Allyl chloride	1.023	0.990	3.2	89	0.00
15 T	Acrylonitrile	0.211	0.221	-4.7	94	0.00
16 T	Acetone	0.204	0.202	1.0	96	0.00
17 T	Carbon Disulfide	1.289	1.157	10.2	81	0.00
18 T	Methyl Acetate	0.591	0.540	8.6	96	0.00
19 T	Methyl tert-butyl Ether	1.389	1.416	-1.9	91	0.00
20 T	Methylene Chloride	0.622	0.566	9.0	90	0.00
21 T	trans-1,2-Dichloroethene	0.523	0.495	5.4	88	0.00
22 T	Diisopropyl ether	2.033	2.065	-1.6	92	0.00
23 T	Vinyl Acetate	1.332	1.425	-7.0	91	0.00
24 P	1,1-Dichloroethane	1.083	1.062	1.9	93	0.00
25 T	2-Butanone	0.261	0.279	-6.9	94	0.00
26 T	2,2-Dichloropropane	0.748	0.722	3.5	90	0.00
27 T	cis-1,2-Dichloroethene	0.595	0.594	0.2	91	0.00
28 T	Bromochloromethane	0.546	0.527	3.5	91	0.00
29 T	Tetrahydrofuran	0.167	0.181	-8.4	96	0.00
30 C	Chloroform	1.027	1.014	1.3#	93	0.00
31 T	Cyclohexane	1.029	0.911	11.5	85	0.00
32 T	1,1,1-Trichloroethane	0.833	0.812	2.5	91	0.00
33 S	1,2-Dichloroethane-d4	0.654	0.670	-2.4	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	91	0.00
35 S	Dibromofluoromethane	0.328	0.334	-1.8	95	0.00
36 T	1,1-Dichloropropene	0.479	0.453	5.4	90	0.00
37 T	Ethyl Acetate	0.367	0.375	-2.2	97	0.00
38 T	Carbon Tetrachloride	0.462	0.434	6.1	90	0.00
39 T	Methylcyclohexane	0.509	0.455	10.6	82	0.00
40 TM	Benzene	1.412	1.381	2.2	90	0.00
41 T	Methacrylonitrile	0.195	0.208	-6.7	93	0.00
42 TM	1,2-Dichloroethane	0.500	0.487	2.6	92	0.00
43 T	Isopropyl Acetate	0.590	0.638	-8.1	95	0.00
44 TM	Trichloroethene	0.360	0.339	5.8	89	0.00
45 C	1,2-Dichloropropane	0.407	0.400	1.7#	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN050119\
 Data File : VN055334.D
 Acq On : 1 May 2019 17:36
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: May 02 04:08:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:16:52 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.223	0.220	1.3	92	0.00
47 T	Bromodichloromethane	0.477	0.482	-1.0	93	0.00
48 T	Methyl methacrylate	0.290	0.296	-2.1	90	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	97	0.00
50 S	Toluene-d8	1.185	1.209	-2.0	94	0.00
51 T	4-Methyl-2-Pentanone	0.332	0.360	-8.4	96	0.00
52 CM	Toluene	0.810	0.808	0.2#	91	0.00
53 T	t-1,3-Dichloropropene	0.415	0.455	-9.6	93	0.00
54 T	cis-1,3-Dichloropropene	0.517	0.540	-4.4	89	0.00
55 T	1,1,2-Trichloroethane	0.301	0.302	-0.3	92	0.00
56 T	Ethyl methacrylate	0.370	0.404	-9.2	92	0.00
57 T	1,3-Dichloropropane	0.528	0.534	-1.1	92	0.00
58 T	2-Chloroethyl Vinyl ether	0.097	0.106	-9.3	91	0.00
59 T	2-Hexanone	0.208	0.225	-8.2	96	0.00
60 T	Dibromochloromethane	0.307	0.325	-5.9	91	0.00
61 T	1,2-Dibromoethane	0.278	0.289	-4.0	92	0.00
62 S	4-Bromofluorobenzene	0.370	0.400	-8.1	93	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	92	0.00
64 T	Tetrachloroethene	0.409	0.368	10.0	89	0.00
65 PM	Chlorobenzene	0.978	0.944	3.5	90	0.00
66 T	1,1,1,2-Tetrachloroethane	0.383	0.372	2.9	92	0.00
67 C	Ethyl Benzene	1.765	1.718	2.7#	89	0.00
68 T	m/p-Xylenes	0.630	0.633	-0.5	91	0.00
69 T	o-Xylene	0.625	0.624	0.2	92	0.00
70 T	Styrene	0.959	1.034	-7.8	92	0.00
71 P	Bromoform	0.212	0.235	-10.8	94	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
73 T	Isopropylbenzene	4.997	4.119	17.6	89	0.00
74 T	N-amyl acetate	1.283	1.390	-8.3	97	0.00
75 P	1,1,2,2-Tetrachloroethane	1.268	1.091	14.0	95	0.00
76 T	1,2,3-Trichloropropane	1.139	1.079	5.3	101	0.00
77 T	Bromobenzene	1.134	1.018	10.2	93	0.00
78 T	n-propylbenzene	5.136	4.524	11.9	87	0.00
79 T	2-Chlorotoluene	3.285	2.793	15.0	89	0.00
80 T	1,3,5-Trimethylbenzene	3.920	3.397	13.3	89	0.00
81 T	trans-1,4-Dichloro-2-butene	0.264	0.257	2.7	92	0.00
82 T	4-Chlorotoluene	2.986	2.702	9.5	90	0.00
83 T	tert-Butylbenzene	3.508	2.816	19.7	88	0.00
84 T	1,2,4-Trimethylbenzene	3.506	3.262	7.0	88	0.00
85 T	sec-Butylbenzene	4.260	3.664	14.0	87	0.00
86 T	p-Isopropyltoluene	3.502	3.161	9.7	86	0.00
87 T	1,3-Dichlorobenzene	1.673	1.569	6.2	89	0.00
88 T	1,4-Dichlorobenzene	1.576	1.525	3.2	90	0.00
89 T	n-Butylbenzene	2.488	2.452	1.4	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN050119\
Data File : VN055334.D
Acq On : 1 May 2019 17:36
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: May 02 04:08:39 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.662	0.543	18.0	86	0.00
91 T	1,2-Dichlorobenzene	1.658	1.565	5.6	90	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.132	0.149	-12.9	95	0.00
93 T	1,2,4-Trichlorobenzene	0.521	0.570	-9.4	84	0.00
94 T	Hexachlorobutadiene	0.617	0.529	14.3	90	0.00
95 T	Naphthalene	1.122	1.180	-5.2	84	0.00
96 T	1,2,3-Trichlorobenzene	0.533	0.576	-8.1	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6