

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050120\
 Data File : VN061230.D
 Acq On : 1 May 2020 15:08
 Operator : JC/MD
 Sample : L2468-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB180-GW-768-770

Quant Time: May 04 03:51:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 29 08:14:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	116738	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	207054	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	198198	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	87852	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	90677	49.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.18%	
35) Dibromofluoromethane	7.55	113	60911	49.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.50%	
50) Toluene-d8	10.06	98	257519	51.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.20%	
62) 4-Bromofluorobenzene	12.38	95	95503	50.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.70%	
Target Compounds						
12) 1,1-Dichloroethene	3.69	96	1323	1.074	ug/l #	64
16) Acetone	3.78	43	7015	4.392	ug/l	92
30) Chloroform	7.33	83	1309	0.467	ug/l	99
44) Trichloroethene	8.80	130	77323	46.320	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050120\
Data File : VN061230.D
Acq On : 1 May 2020 15:08
Operator : JC/MD
Sample : L2468-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB180-GW-768-770

Quant Time: May 04 03:51:39 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 29 08:14:26 2020
Response via : Initial Calibration

