

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050120\
 Data File : VN061247.D
 Acq On : 2 May 2020 2:04
 Operator : JC/MD
 Sample : VN0501WBL02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0501WBL02

Quant Time: May 04 07:24:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 29 08:14:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	100789	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	179286	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	169402	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	71885	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.98	65	79715	49.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.96%	
35) Dibromofluoromethane	7.54	113	51872	48.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.86%	
50) Toluene-d8	10.06	98	224220	51.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.76%	
62) 4-Bromofluorobenzene	12.37	95	80198	49.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.62%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN050120\
Data File : VN061247.D
Acq On : 2 May 2020 2:04
Operator : JC/MD
Sample : VN0501WBL02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0501WBL02

Quant Time: May 04 07:24:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 29 08:14:26 2020
Response via : Initial Calibration

