

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050120\
 Data File : VN061250.D
 Acq On : 2 May 2020 4:15
 Operator : JC/MD
 Sample : PB128586TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB128586TB

Quant Time: May 04 07:40:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 29 08:14:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	102319	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.54	114	183004	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	177571	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	73944	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.98	65	81274	50.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.40%	
35) Dibromofluoromethane	7.54	113	40630	37.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.10%	
50) Toluene-d8	10.06	98	235273	53.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.66%	
62) 4-Bromofluorobenzene	12.37	95	84731	51.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.08%	
Target Compounds						
16) Acetone	3.77	43	8662	8.497	ug/l	96
43) Isopropyl Acetate	8.12	43	44379	12.451	ug/l #	90
80) 1,3,5-Trimethylbenzene	12.71	105	4552	1.012	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050120\
Data File : VN061250.D
Acq On : 2 May 2020 4:15
Operator : JC/MD
Sample : PB128586TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB128586TB

Quant Time: May 04 07:40:28 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 29 08:14:26 2020
Response via : Initial Calibration

