

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050120\
 Data File : VN061251.D
 Acq On : 2 May 2020 4:39
 Operator : JC/MD
 Sample : L2122-15
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU-03-043020

Quant Time: May 04 07:43:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 29 08:14:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	106461	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	193921	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	189607	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	81705	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	87011	51.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.30%	
35) Dibromofluoromethane	7.54	113	55002	47.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.94%	
50) Toluene-d8	10.06	98	247071	52.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.70%	
62) 4-Bromofluorobenzene	12.37	95	90708	51.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.12%	
Target Compounds						
16) Acetone	3.78	43	9435	9.160	ug/l	99
20) Methylene Chloride	4.50	84	1528	1.163	ug/l	96
43) Isopropyl Acetate	8.12	43	41146	10.894	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050120\
Data File : VN061251.D
Acq On : 2 May 2020 4:39
Operator : JC/MD
Sample : L2122-15
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU-03-043020

Quant Time: May 04 07:43:30 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 29 08:14:26 2020
Response via : Initial Calibration

