

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050120\
 Data File : VN061255.D
 Acq On : 2 May 2020 6:12
 Operator : JC/MD
 Sample : L2452-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 209

Quant Time: May 04 07:58:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 29 08:14:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	109893	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	202920	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	204086	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	88094	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	91601	52.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.36%	
35) Dibromofluoromethane	7.55	113	55403	46.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.34%	
50) Toluene-d8	10.06	98	261749	53.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.02%	
62) 4-Bromofluorobenzene	12.38	95	96686	52.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.06%	
Target Compounds						
16) Acetone	3.77	43	20925	26.169	ug/l	96
20) Methylene Chloride	4.50	84	2301	1.697	ug/l	88
43) Isopropyl Acetate	8.13	43	37373	9.456	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050120\
Data File : VN061255.D
Acq On : 2 May 2020 6:12
Operator : JC/MD
Sample : L2452-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
209

Quant Time: May 04 07:58:00 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 29 08:14:26 2020
Response via : Initial Calibration

