

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050120\
 Data File : VN061260.D
 Acq On : 2 May 2020 8:08
 Operator : JC/MD
 Sample : L2463-22
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JHDS-SOUTH-C

Quant Time: May 04 08:27:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 29 08:14:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	94360	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	174789	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	176372	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	75891	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	81386	54.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.02%	
35) Dibromofluoromethane	7.55	113	49569	47.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.92%	
50) Toluene-d8	10.06	98	226995	53.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.74%	
62) 4-Bromofluorobenzene	12.37	95	83078	52.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.80%	
Target Compounds						
16) Acetone	3.78	43	12596	16.657	ug/l	94
20) Methylene Chloride	4.50	84	1324	1.137	ug/l #	83
43) Isopropyl Acetate	8.12	43	38359	11.268	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050120\
Data File : VN061260.D
Acq On : 2 May 2020 8:08
Operator : JC/MD
Sample : L2463-22
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JHDS-SOUTH-C

Quant Time: May 04 08:27:58 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 29 08:14:26 2020
Response via : Initial Calibration

