

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050120\
 Data File : VN061263.D
 Acq On : 2 May 2020 9:18
 Operator : JC/MD
 Sample : L2467-11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SS040RW056-200429

Quant Time: May 04 08:37:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 29 08:14:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	100801	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	182852	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	174533	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	76490	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	83264	52.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.40%	
35) Dibromofluoromethane	7.55	113	50938	47.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.22%	
50) Toluene-d8	10.06	98	229022	51.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.92%	
62) 4-Bromofluorobenzene	12.38	95	84213	50.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.54%	
Target Compounds						
4) Vinyl Chloride	2.16	62	25426	17.292	ug/l	98
27) cis-1,2-Dichloroethene	6.78	96	156562	115.201	ug/l	90
44) Trichloroethene	8.80	130	167427	113.572	ug/l	97
64) Tetrachloroethene	10.60	164	42717	28.259	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN050120\
Data File : VN061263.D
Acq On : 2 May 2020 9:18
Operator : JC/MD
Sample : L2467-11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SS040RW056-200429

Quant Time: May 04 08:37:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 29 08:14:26 2020
Response via : Initial Calibration

