

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050120\
 Data File : VN061265.D
 Acq On : 2 May 2020 10:04
 Operator : JC/MD
 Sample : L2467-09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Z3-DUP-0548-200428

Quant Time: May 04 08:49:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 29 08:14:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	100265	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	177795	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	170197	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	75391	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	81010	51.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.12%	
35) Dibromofluoromethane	7.55	113	34225	32.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	65.10%	
50) Toluene-d8	10.06	98	221103	51.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.18%	
62) 4-Bromofluorobenzene	12.38	95	79919	49.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.10%	
Target Compounds						
4) Vinyl Chloride	2.15	62	3814458	2608.029	ug/l #	77
12) 1,1-Dichloroethene	3.70	96	12745	12.048	ug/l	90
21) trans-1,2-Dichloroethene	4.99	96	69679	60.489	ug/l	98
27) cis-1,2-Dichloroethene	6.78	96	6422877	4751.317	ug/l	81
39) Methylcyclohexane	9.05	83	4441	2.225	ug/l #	76
52) Toluene	10.12	92	2606	0.816	ug/l	100
65) Chlorobenzene	11.41	112	11120	3.294	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN050120\
Data File : VN061265.D
Acq On : 2 May 2020 10:04
Operator : JC/MD
Sample : L2467-09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Z3-DUP-0548-200428

Quant Time: May 04 08:49:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 29 08:14:26 2020
Response via : Initial Calibration

