


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN050120\
 Data File : VN061219.D
 Acq On : 1 May 2020 9:35
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Title : SW846 8260
 Last Update : Wed Apr 29 08:14:26 2020



AutoFind: Scans 3353, 3354, 3355; Background Corrected with Scan 3341

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.5	6261	PASS
75	95	30	60	53.5	14924	PASS
95	95	100	100	100.0	27877	PASS
96	95	5	9	6.8	1893	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	74.3	20714	PASS
175	174	5	9	7.1	1467	PASS
176	174	95	101	96.2	19936	PASS
177	176	5	9	6.3	1252	PASS