

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050123\
 Data File : VN077540.D
 Acq On : 01 May 2023 12:17
 Operator : JC\MD
 Sample : VN0501WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0501WBL01

Quant Time: May 01 22:00:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	610700	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1154410	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	1042266	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	367500	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	415355	52.122	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.240%
35) Dibromofluoromethane	8.172	113	358061	51.605	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.200%
50) Toluene-d8	10.572	98	1402605	52.745	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.480%
62) 4-Bromofluorobenzene	12.854	95	430896	49.526	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050123\
Data File : VN077540.D
Acq On : 01 May 2023 12:17
Operator : JC\MD
Sample : VN0501WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0501WBL01

Quant Time: May 01 22:00:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

