

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050123\
 Data File : VN077546.D
 Acq On : 01 May 2023 16:48
 Operator : JC\MD
 Sample : 02547-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-TB-20230426

Quant Time: May 01 22:04:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	613033	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1138747	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	996960	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	351564	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	400358	50.049	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.100%
35) Dibromofluoromethane	8.172	113	354865	51.848	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.700%
50) Toluene-d8	10.566	98	1361815	51.915	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.840%
62) 4-Bromofluorobenzene	12.854	95	422901	49.276	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050123\
Data File : VN077546.D
Acq On : 01 May 2023 16:48
Operator : JC\MD
Sample : 02547-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-TB-20230426

Quant Time: May 01 22:04:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

