

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050124\
 Data File : VN081946.D
 Acq On : 01 May 2024 10:51
 Operator : JC\MD
 Sample : VN0501MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0501MBL01

Quant Time: May 02 01:27:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	188892	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	379904	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	327096	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	126968	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	144245	45.246	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.500%
35) Dibromofluoromethane	8.165	113	108330	47.361	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.720%
50) Toluene-d8	10.565	98	462521	51.486	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.980%
62) 4-Bromofluorobenzene	12.853	95	155954	43.916	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	87.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050124\
Data File : VN081946.D
Acq On : 01 May 2024 10:51
Operator : JC\MD
Sample : VN0501MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0501MBL01

Quant Time: May 02 01:27:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration

