

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050124\
 Data File : VN081955.D
 Acq On : 01 May 2024 14:38
 Operator : JC\MD
 Sample : P2306-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-4-20240429-A

Quant Time: May 02 01:37:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	193350	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	381568	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	333146	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	134570	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	148382	45.471	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.940%
35) Dibromofluoromethane	8.165	113	112285	48.876	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.760%
50) Toluene-d8	10.565	98	469702	52.057	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.120%
62) 4-Bromofluorobenzene	12.853	95	158529	44.446	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	88.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050124\
Data File : VN081955.D
Acq On : 01 May 2024 14:38
Operator : JC\MD
Sample : P2306-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-4-20240429-A

Quant Time: May 02 01:37:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration

