

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050124\
 Data File : VN081959.D
 Acq On : 01 May 2024 16:15
 Operator : JC\MD
 Sample : P2329-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-18B-64-043024

Quant Time: May 02 01:43:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	192190	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	386265	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	336703	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	134370	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	147139	45.362	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	90.720%		
35) Dibromofluoromethane	8.165	113	111261	47.841	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.680%		
50) Toluene-d8	10.571	98	466910	51.119	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	102.240%		
62) 4-Bromofluorobenzene	12.853	95	160053	44.328	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery =	88.660%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.083	85	29114	10.791	ug/l	95
4) Vinyl Chloride	2.512	62	11579	4.545	ug/l	94
9) 1,1,2-Trichlorotrifluo...	4.377	101	336242	146.953	ug/l	92
12) 1,1-Dichloroethene	4.342	96	88727	38.117	ug/l	98
21) trans-1,2-Dichloroethene	5.794	96	20591	8.113	ug/l #	91
24) 1,1-Dichloroethane	6.571	63	78672	15.864	ug/l	97
27) cis-1,2-Dichloroethene	7.488	96	1413683	476.668	ug/l	99
40) Benzene	8.606	78	19066	1.679	ug/l	91
42) 1,2-Dichloroethane	8.671	62	12600	2.987	ug/l	99
44) Trichloroethene	9.359	130	12526193	4783.450	ug/l	70
64) Tetrachloroethene	11.106	164	139422	57.013	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050124\
Data File : VN081959.D
Acq On : 01 May 2024 16:15
Operator : JC\MD
Sample : P2329-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-18B-64-043024

Quant Time: May 02 01:43:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration

