

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050124\
 Data File : VN081963.D
 Acq On : 01 May 2024 17:51
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: May 02 01:49:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	172486	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	346961	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	300254	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	115428	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	137031	47.072	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.140%
35) Dibromofluoromethane	8.171	113	101926	48.792	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.580%
50) Toluene-d8	10.571	98	432234	52.683	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.360%
62) 4-Bromofluorobenzene	12.847	95	142663	43.988	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	87.980%
Target Compounds						
					Qvalue	
44) Trichloroethene	9.353	130	6478	2.754	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050124\
Data File : VN081963.D
Acq On : 01 May 2024 17:51
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: May 02 01:49:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration

