

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN050219\
 Data File : VN055346.D
 Acq On : 2 May 2019 13:23
 Operator : JC/SP
 Sample : K2658-17
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EP1-SB-EB-01

Quant Time: May 03 03:42:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:16:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	357366	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	609712	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	543345	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	153060	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	244624	52.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.64%	
35) Dibromofluoromethane	7.59	113	195571	48.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.74%	
50) Toluene-d8	10.09	98	744124	51.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.98%	
62) 4-Bromofluorobenzene	12.40	95	228230	50.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.06%	
Target Compounds						
3) Chloromethane	2.06	50	8502	1.568	ug/l	93
16) Acetone	3.83	43	20481	14.060	ug/l	94
84) 1,2,4-Trimethylbenzene	13.04	105	35161	3.276	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN050219\
Data File : VN055346.D
Acq On : 2 May 2019 13:23
Operator : JC/SP
Sample : K2658-17
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EP1-SB-EB-01

Quant Time: May 03 03:42:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration

